

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME **CYTOMEGALOVIRUS IgM-CHORUS**

Code	81011		
Lot No.	026	2025-07	2024-04

REAGENTS

Label	Description	Lot	Exp.	NO
D2171	Cytomegalovirus IgM	026	2025-07	6X6
B2171	CALIBRATORE Cytomegalovirus IgM	026	2025-07	1
C2171	CONTROL SERUM Cytomegalovirus IgM	026	2025-07	1
G2171	CONFIRMATORY TEST	026	2025-07	5X1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	572	≥ 100
Control Serum (INDEX)	1,9	1,5-2,8
Negative Control (INDEX)	0,2	<0,7

CALIBRATOR

0	7	9	0	2	6	7	0	0	0	0	0	4	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Ateneu*

079026700000470000
Cytomegalovirus IgM
CALIBRATOR 0,175 mL IVD
LOT 026
2025-07 2°C-8°C
DIESSE Monteriggioni Italy CE 0123 B2171

CONTROL SERUM

0	7	9	0	2	6	7	1	5	2	8	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Ateneu*

079026715280000001
Cytomegalovirus IgM
CONTROL 0,425 mL IVD
LOT 026
2025-07 2°C-8°C
DIESSE Monteriggioni Italy CE 0123 C2171

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Ateneu*

DATE *25-06-2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.