

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554

**DIESSE**

QUALITY CONTROL CERTIFICATE

KIT NAME CHLAMYDOPHILA PNEUMONIAE IgM - CHORUS

Code 81251

Lot No. 027

2025-04

2024-06

REAGENTS

| Label | Description                         | Lot. | Exp.    | NO  |
|-------|-------------------------------------|------|---------|-----|
| D8600 | Chlamydomphila pneumoniae IgM       | 027  | 2025-04 | 6X6 |
| B8600 | CALIBRATORE Chlamydomphila p. IgM   | 027  | 2025-04 | 1   |
| C8600 | CONTROL SERUM Chlamydomphila p. IgM | 027  | 2025-04 | 1   |

CONTROLS

|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                          |     |            |
|--------------------------|-----|------------|
| Calibrator (mOD)         | 690 | $\geq 100$ |
| Control Serum (Index)    | 1,9 | 1,2-2,3    |
| Negative Control (Index) | 0,2 | $\leq 0.7$ |


CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 3 | 0 | 0 | 2 | 7 | 6 | 0 | 0 | 0 | 0 | 0 | 5 | 8 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES ☒

Operator Signature .....



CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 3 | 0 | 0 | 2 | 7 | 6 | 1 | 2 | 2 | 3 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES ☒

Operator Signature .....



This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 26.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.