



EU DECLARATION OF CONFORMITY

MANUFACTURER:	DIESESE DIAGNOSTICA SENESE SPA STRADA DEI LAGHI 39 53035 MONTERIGGIONI (SI), ITALY
SINGLE REGISTRATION NUMBER	IT-MF-000013311
AUTHORIZED REPRESENTATIVE:	NOT APPLICABLE
PRODUCT: CODE:	CHORUS ZIKA IgG 81158
INTENDED PURPOSE:	<p>CHORUS ZIKA IgG is an immunoassay kit for the semi-quantitative automated determination of IgG class anti-Zika virus antibodies.</p> <p>The test is performed on human serum using a disposable device applied on CHORUS/ CHORUS TRIO/ CHORUS EVO instruments.</p> <p>The kit is intended to detect exposure to Zika virus infection as an aid to related diagnosis.</p> <p>It is to be used only by professional laboratory personnel.</p>
BASIC UDI-DI	803389132CHORUSZIG00D3
UDI-DI	08033891324131
RISK CLASS:	CLASS C
CLASSIFICATION RULE:	RULE 3e
CONFORMITY ASSESSMENT ROUTE:	ANNEX IX
<p>WE HEREWITH DECLARE THAT THE ABOVE MENTIONED PRODUCT IS IN CONFORMITY WITH REGULATION (EU) 2017/746 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL ON IN VITRO DIAGNOSTIC MEDICAL DEVICE. ALL SUPPORTING DOCUMENTATION IS RETAINED UNDER THE PREMISES OF THE MANUFACTURER.</p>	
REFERENCE TO ANY CS APPLIED:	NOT APPLICABLE
NOTIFIED BODY:	TÜV SÜD PRODUCT SERVICE GMBH RIDLERSTAßE 65 – 80339 MUNICH GERMANY No. 0123

(EU) CERTIFICATE:

V12 056726 0006 Rev. 02

REVISION:

0

PLACE, DATE OF ISSUE:

MONTERIGGIONI, 27 JUNE 2024

EXPIRY DATE:

2027-06-26

THE PRESENT EU DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE AND EXCLUSIVE RESPONSIBILITY OF THE MANUFACTURER.

SIGNATURE:



CHIARA MUZZI
PERSON RESPONSIBLE FOR THE REGULATORY
COMPLIANCE

THIS IS THE CERTIFIED COPY OF THE ORIGINAL DOCUMENT STORED IN ARCHIVE OF DIESSE DIAGNOSTICA SENESE SPA

ISSUED: MONTERIGGIONI, 2024-06-27



MAGDALENA STOCZKO
REGULATORY SUPERVISOR