

DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1 IgG RECOMBINANT - CHORUS

Code	81023		
Lot No.	042	2025-06	2024-06

REAGENTS

Label	Description	Lot	Exp.	NO
D8681	Herpes simplex 1 IgG Recombinant	042	2025-06	6x6
B8681	CALIBRATORE Herpes simplex 1 IgG Recombinant	042	2025-06	1
C8681	CONTROL SERUM Herpes simplex 1 IgG Recombinant	042	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	398	> 100
CONTROL SERUM (INDEX)	2,0	1,5-2,7
NEGATIVE CONTROL (INDEX)	0,0	<0,7

CALIBRATOR

2	5	8	0	4	2	6	0	0	0	0	0	4	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Abun*

25804260000490000
Herpes simplex 1 IgG Recombinant
CALIBRATOR 0.175 mL IVD
LOT 042
2025-06
DIESSSE Monteriggioni Italy
CE 0123

CONTROL SERUM

2	5	8	0	4	2	6	1	5	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Abun*

258042615270000001
Herpes simplex 1 IgG Recombinant
CONTROL SERUM 0.425 mL IVD
LOT 042
2025-06
DIESSSE Monteriggioni Italy
CE 0123

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Abun*

DATE 28.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.