DIESSE DIAGNOSTICA SENESE SpA



53035 Monteriggioni (SI), Italy, Strada dei Laghi 39 Tel. +39 0577319554

		QUALITY CONT	ROL CERTIFICAT	ΤΕ	1487.45		
KIT NAME C	LEANING SOLUTION	2000		PER.	重新不定		1 gh
Code 83609 Lot No. 119	Ω	2025-12	<u>سا</u> 20:	24-06			
REAGENTS							
Label B2570 (CLE	Description EANING SOLUTION 2000		Lo		Exp.	NO	
D2370 CEL	EANING SOLUTION 2000			119	2025-12	4	
CONTROLS				4000	Park		115
		Found results	Expected results	5			
Density		1,052	1.040 - 1.060]			

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 2266 624